

This product contains propane, oxygen and nitrogen, substances subject to the Pennsylvania Worker and Community Right-To-Know Act.

PRODUCT IDENTITY

LABEL IDENTITY - MSA P/N 455129 Calibration Check Gas, 0.6% Propane in Air
CHEMICAL NAME - Propane, Oxygen, Nitrogen Mixture
ADDITIONAL IDENTITIES - MSA P/N 455129 Calibration Gas
FORMULA - C₃H₈ in Air

APPLICABLE CHEMICAL CONTENTS

	<u>%</u>	<u>TWA</u>
Propane (CAS 74-98-6)	0.6	1000 ppm*
Air	Balance	None

*Propane as an alkane (ACGIH 2009), OSHA TWA=1000 ppm

NOTE: Gas under pressure, 115 PSIG at 70°F, Approx. 3 Liters gas at atmospheric pressure

PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE AND ODOR - Colorless Gas, Faint Odor
BOILING POINT - N/A*
VAPOR PRESSURE - N/A*
VAPOR DENSITY (AIR = 1) - > 1
SOLUBILITY IN WATER - Propane - 6.5 cm³/100 ml (18°C)
Oxygen - 3.2 cm³/100 ml (25°C)
Nitrogen - 2.3 cm³/100 ml (0°C)

SPECIFIC GRAVITY (H₂O = 1) - N/A*
PERCENT VOLATILE BY VOLUME - N/A*

N/A* - Not Applicable

PHYSICAL HAZARD INFORMATION

PHYSICAL HAZARD - Compressed Gas, 115 PSIG at 70°F
CONDITIONS OR MATERIALS TO AVOID - None
FLASH POINT - N/A* LEL - N/A* UEL - N/A*
EXTINGUISHING MEDIA - This Calibration Gas Mixture is not flammable
SPECIAL FIRE FIGHTING PROCEDURES - See Next Item
UNUSUAL FIRE AND EXPLOSION HAZARDS - Gas under pressure, 115 PSIG at 70°F. Do not exceed 120°F.

HEALTH HAZARDS

HEALTH HAZARDS - Propane as an alkane CNS depression, cardiac sensitization.

SIGNS AND SYMPTOMS OF EXPOSURE -. CNS depression and cardiac sensitization (1000 ppm).

PRIMARY ROUTES OF ENTRY - Inhalation

TARGET ORGANS – CNS, and heart

MEDICAL CONDITIONS GENERALLY RECOGNIZED AS BEING AGGRAVATED BY EXPOSURE - No Information

EXPOSURE LIMITS – Propane as an alkane 1000 ppm TWA

CARCINOGENICITY DATA - NIOSH RTECS, OSHA, NTP or IARC does not list component gases.

EMERGENCY AND FIRST AID PROCEDURES - None

SAFE HANDLING AND USE

HYGIENIC PRACTICES - Avoid Breathing Gas

PROTECTIVE MEASURES DURING REPAIR AND MAINTENANCE OF CONTAMINATED EQUIPMENT - N/A

PROCEDURES FOR SPILL OR LEAK CLEANUP - Ventilate area

WASTE DISPOSAL - - Do not puncture or incinerate cylinder. Before discarding cylinder, slowly release contents to a safe exhaust. Dispose of cylinder in accordance with local, state and federal regulations.

STORAGE - Store in a cool, dry, well-ventilated area. Do not exceed 120°F.

CONTROL MEASURES

PERSONAL PROTECTIVE EQUIPMENT - Due to the limited amount of gas in the cylinder, and the low release rate employed in instrument calibration, respiratory protection is not indicated under conditions of intended use.

ENGINEERING CONTROLS - None

WORK PRACTICES - Use in well-ventilated areas. Follow the calibration procedure detailed in the MSA instruction manual provided with the instrument under calibration.

DATE OF PREPARATION - Rev. 12, December 2009

WARNING: This is a hazardous chemical product. By following the directions and warnings provided with this product, the hazards associated with the use of this product can be greatly reduced but never entirely eliminated. Mine Safety Appliances Company makes no warranties, expressed or implied, with respect to this product and EXPRESSLY DISCLAIMS THE WARRANTY OF MERCHANTABILITY AND ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE. Users assume all risks in handling, using or storing this product.