



---

---

**HEALTH HAZARDS**

---

---

HEALTH HAZARDS - None

SIGNS AND SYMPTOMS OF EXPOSURE - N/A\*

PRIMARY ROUTES OF ENTRY - Inhalation

TARGET ORGANS - N/A\*

MEDICAL CONDITIONS GENERALLY RECOGNIZED AS BEING AGGRAVATED BY EXPOSURE - N/A

EXPOSURE LIMITS - N/A\*

CARCINOGENICITY DATA - NIOSH RTECS, OSHA, NTP or IARC does not list component gases.

EMERGENCY AND FIRST AID PROCEDURES – None

N/A\* - Not Applicable

---

---

---

---

**SAFE HANDLING AND USE**

---

---

HYGIENIC PRACTICES - None

PROTECTIVE MEASURES DURING REPAIR AND MAINTENANCE OF CONTAMINATED EQUIPMENT - N/A\*

PROCEDURES FOR SPILL OR LEAK CLEANUP - None

WASTE DISPOSAL - Do not puncture or incinerate cylinder. Before discarding cylinder, slowly release contents to a safe exhaust. Dispose of cylinder in accordance with local, state and federal regulations.

STORAGE - Store in a cool, dry, well-ventilated area. Do not exceed 120°F.

---

---

---

---

**CONTROL MEASURES**

---

---

PERSONAL PROTECTIVE EQUIPMENT - None

ENGINEERING CONTROLS - None

WORK PRACTICES - Follow the calibration procedure detailed in the MSA instruction manual provided with the instrument under calibration.

DATE OF PREPARATION - Rev. 10, September 2011

WARNING: This is a hazardous chemical product. By following the directions and warnings provided with this product, the hazards associated with the use of this product can be greatly reduced but never entirely eliminated. Mine Safety Appliances Company makes no warranties, expressed or implied, with respect to this product and EXPRESSLY DISCLAIMS THE WARRANTY OF MERCHANTABILITY AND ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE. Users assume all risks in handling, using or storing this product.