

This product contains carbon monoxide, methane, hydrogen sulfide, oxygen and nitrogen, substances subject to the Pennsylvania Worker and Community Right-To-Know Act.

PRODUCT IDENTITY

LABEL IDENTITY - MSA P/N 814559 Calibration Check Gas, 300 ppm Carbon Monoxide, 1.25% Methane, 15% Oxygen, 25 ppm Hydrogen Sulfide, Balance Nitrogen

CHEMICAL NAME - Carbon monoxide, methane, oxygen, hydrogen sulfide, nitrogen

ADDITIONAL IDENTITIES - MSA P/N 814559 calibration gas, MSA P/N 814559 Squirt™ gas

FORMULA - $CO + CH_4 + O_2 + H_2S + N_2$

APPLICABLE CHEMICAL CONTENTS

	<u>%</u>	<u>TWA</u>	<u>STEL</u>
Carbon Monoxide (CAS 630-08-0) (ACGIH 2010)	0.03	25 ppm	
Methane (CAS 74-82-8)	1.25	1000 ppm*	
Oxygen (CAS 7782-44-7)	15	None	
Hydrogen Sulfide (CAS 7783-06-4) (ACGIH 2010)	0.0025	1 ppm	5 ppm
Nitrogen (CAS 7727-37-9)	Balance	None	

*Methane as an alkane (ACGIH 2010)

NOTE: Gas Under Pressure, 155 PSIG at 70°F, Approx. 11 Liters Gas at Atmospheric Pressure

PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE AND ODOR - Colorless, malodorous gas

Following information is for Nitrogen the main component of this gas mixture

BOILING POINT - -320.4°F (-195.8 °C) SPECIFIC GRAVITY (air = 1) @70°F (21.1 °C): 0.906

VAPOR PRESSURE @70°F (21.1 °C): N/A* PERCENT VOLATILE BY VOLUME - N/A*

GAS DENSITY @32°F (0 °C) and 1 atm: 0.072 lbs/ft³ (1.153 kg/ m³)

VAPOR DENSITY (AIR = 1) - Approx. 1

SOLUBILITY IN WATER - Carbon Monoxide - 3.5 cm³/100 ml (0°C)
Methane - 9 cm³/100 ml (20°C)
Oxygen - 3.2 cm³/100 ml (25°C)
Hydrogen Sulfide - 437 cm³/100 ml (0°C)
Nitrogen - 2.3 cm³/100 ml (0°C)

* N/A - Not Applicable

PHYSICAL HAZARD INFORMATION

PHYSICAL HAZARD - Compressed gas, 155 PSIG at 70°F

CONDITIONS OR MATERIALS TO AVOID - None

FLASH POINT - N/A LEL - N/A UEL - N/A

EXTINGUISHING MEDIA - This gas mixture is not flammable.

SPECIAL FIRE FIGHTING PROCEDURES - See next item

UNUSUAL FIRE AND EXPLOSION HAZARDS - Gas under pressure, 155 PSIG at 70°F. Do not exceed 120°F.

HEALTH HAZARDS

HEALTH HAZARDS –

- Methane as an alkane CNS depression and cardiac sensitization at 1000 ppm.
- Carbon monoxide TC_{LO} (human) is reportedly 650 ppm/45 minutes; LC_{LO} (human) 5000 ppm/5 minutes.
- Hydrogen sulfide LC_{LO} (human) 800 ppm/5 minutes. No effects known below 10 ppm.

SIGNS AND SYMPTOMS OF EXPOSURE –

- Methane: Cardiac sensitization and Central nervous system depression at 1000ppm..
- Carbon Monoxide exposure to 500-1000 ppm CO may cause headache, rapid breathing, nausea, weakness, dizziness and confusion.
- Hydrogen sulfide exposure is an irritant of the eyes and respiratory tract.

PRIMARY ROUTES OF ENTRY - Inhalation, eyes**TARGET ORGANS –**

- Methane affects the central nervous system, and heart.
- Carbon monoxide: Lungs, Blood, Tissues. Carbon monoxide at toxic concentrations causes tissue hypoxia (lack of oxygen) by preventing blood from transporting sufficient oxygen.
- Hydrogen sulfide is an irritant of the eyes and respiratory tract.

MEDICAL CONDITIONS GENERALLY RECOGNIZED AS BEING AGGRAVATED BY EXPOSURE - Carbon monoxide burden may aggravate angina pectoris. Pregnant women are reportedly more sensitive to Carbon monoxide than others. Effects of Carbon monoxide exposure are aggravated by heavy labor, heat stress and high altitude.

EXPOSURE LIMITS –

- Methane as an alkane 1000 ppm. (ACGIH 2010).
- Carbon Monoxide 25 ppm (ACGIH 2010). OSHA TWA 35 ppm
- Hydrogen sulfide: TWA is 1 ppm, STEL 5 ppm. (ACGIH 2010).

CARCINOGENICITY DATA - NIOSH RTECS, OSHA, NTP or IARC does not list component gases.

EMERGENCY AND FIRST AID PROCEDURES - Remove from exposure. Administer oxygen. Consult physician immediately.

SAFE HANDLING AND USE

HYGIENIC PRACTICES - Avoid breathing gas.

PROTECTIVE MEASURES DURING REPAIR AND MAINTENANCE OF CONTAMINATED EQUIPMENT - N/A

PROCEDURES FOR SPILL OR LEAK CLEANUP - Ventilate area. Avoid breathing gas.

WASTE DISPOSAL - Do not puncture or incinerate cylinder. Before discarding cylinder, slowly release contents to a safe exhaust. Dispose of cylinder in accordance with local, state, and federal regulations.

STORAGE - Store in a cool, dry, well-ventilated area. Do not exceed 120°F.

CONTROL MEASURES

PERSONAL PROTECTIVE EQUIPMENT - Due to the limited amount of gas in the cylinder, and the low release rate employed in instrument calibration, respiratory protection is not indicated under conditions of intended use.

ENGINEERING CONTROLS - Mechanical ventilation is suitable.

WORK PRACTICES - Avoid breathing gas. Use in well-ventilated areas. Follow the calibration procedure detailed in the MSA instruction manual provided with the instrument under calibration.

DATE OF PREPARATION - Rev. 7, October 2010

WARNING: This is a hazardous chemical product. By following the directions and warnings provided with this product, the hazards associated with the use of this product can be greatly reduced but never entirely eliminated. Mine Safety Appliances Company makes no warranties, expressed or implied, with respect to this product and EXPRESSLY DISCLAIMS THE WARRANTY OF MERCHANTABILITY AND ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE. Users assume all risks in handling, using or storing this product.